



Heatable dilution system up to 150 °C for large droplets up to 10 μm

Description

The dilution of large droplets is particularly important when measuring highly concentrated droplet aerosols. Since large droplets are difficult to dilute, standard systems only work up to a size of 1 - 2 μm . The model variant LDD 100 H can be heated up to 150 °C and thus prevents condensation.

The newly developed dilution system LDD 100 H (dilution factor 100) is the first system to dilute almost loss-free large droplets up to 10 μm .

LDD 100:

Particle size	Number counts without dilution	Number counts with dilution	Dilution factor
5 μm	304322	3043	100.01
7 μm	236687	2370	99.87

Chart 1: Dilution of monodisperse DEHS droplets with LDD 100




Fig. 1: Distribution of LDD 100 (7 μm)

Benefits

- Defined dilution of large droplets of factor 100
- Proven dilution factor 100 for droplet sizes up to 7 μm
- Easy connection with Promo® and welas® digital aerosol spectrometers
- Internal pump for autonomous operations
- Resistant to pressure fluctuations of ± 200 mbar
- Simple handling
- Robust, durable, low maintenance
- Cost effective

Datasheet

<i>Parameter</i>	<i>Description</i>
Power consumption	200 W
Dimensions	Steuereinheit: 185 • 450 • 315 mm (H • B • T) Verdünnungseinheit: 250 • 145 • 120 mm (H • B • T)
Weight	Steuereinheit: 10,2 kg Verdünnungseinheit: 2,9 kg
Installation conditions	0 – 40 °C
Dilution factor	1 : 10 1 : 100
Special features	Heatable up to 150 °C
Volume flow (suction flow)	0.5 l/min $\hat{=}$ 0.03 m ³ /h

Applications

- Measurement of blow-by aerosols according to ISO 17536
- Dilution of compressed air
- Measurement of cooling lubricant aerosols

Palas GmbH
Partikel- und Lasermesstechnik
Greschbachstrasse 3 b
76229 Karlsruhe
Germany

Managing Partner:
Dr.-Ing. Maximilian Weiß
Commercial Register:
register court: Mannheim
company registration number: HRB 103813
USt-Id: DE143585902



Contact: E-Mail: mail@palas.de Internet: www.palas.de Tel: +49 (0)721 96213-0 Fax: +49 (0)721 96213-33