



Heatable sampling tube

Description

During the sample taking of hot aerosols, particle condensation from the gas phase may occur if the aerosol cools down. To avoid that effect, the whole sample taking unit including the aerosol sensors, sampling lines and dilution systems must be heated. The sample taking should be isothermal to avoid changes of the particle concentration and the particle size distribution. The heatable sampling tubes PS 1000 / PS 2000 and PS 3000 can be heated up to 120 °C and can be adapted to the heated welas® aerosol sensors easily. The heating control is part of the delivery content. For heating temperatures up to 250 °C Palas^{small>®</small>} offers the variation with the articles PS 1000 HT / PS 2000 HT and PS 3000 HT.

Benefits

- Actively heated, flexible sampling tube
- Easy use with welas® aerosol sensors
- Avoiding of condensation effects during aerosol sampling by heating to the appropriate temperature
- Selection for temperatures up to 120 °C or up to 250 °C

Datasheet

<i>Parameter</i>	<i>Description</i>
Power supply	115 V or 250 V
Dimensions	1 m, 2 m or 3 m (length)
Temperature regulation	Heatable to 120 °C or 250 °C
Diameter	6 mm (inside), 8 mm (outside)

Applications

- Isothermal sample taking of hot aerosols
- Blow-by measurements

Palas GmbH
Partikel- und Lasermesstechnik
Greschbachstrasse 3 b
76229 Karlsruhe
Germany

Managing Partner:
Dr.-Ing. Maximilian Weiß
Commercial Register:
register court: Mannheim
company registration number: HRB 103813
USt-Id: DE143585902



Contact: E-Mail: mail@palas.de Internet: www.palas.de Tel: +49 (0)721 96213-0 Fax: +49 (0)721 96213-33